

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.007 °C	11.6 °C	1.613 °C	± 0.53 °C	± 4.0E-01 °C
20 °C	20.035 °C	21.7 °C	1.625 °C	± 0.53 °C	± 4.0E-01 °C
30 °C	30.163 °C	31.8 °C	1.607 °C	± 0.53 °C	± 4.0E-01 °C
40 °C	40.291 °C	41.8 °C	1.509 °C	± 0.53 °C	± 4.0E-01 °C