

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	9.787 °C	10.2 °C	0.413 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.927 °C	32.3 °C	0.373 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	34.537 %HR	38.0 %HR	3.463 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.059 %HR	72.2 %HR	-2.859 %HR	± 3.0 %HR	± 1.9E+00 %HR