

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.408 °C	10.6 °C	0.192 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.619 °C	32.8 °C	0.181 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.580 %HR	37.7 %HR	2.120 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	76.7 %HR	1.140 %HR	± 3.0 %HR	± 1.9E+00 %HR