

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	9.787 °C	10.2 °C	0.413 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.927 °C	32.4 °C	0.473 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	34.537 %HR	37.4 %HR	2.863 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.059 %HR	73.5 %HR	-1.559 %HR	± 3.0 %HR	± 1.9E+00 %HR