

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -30 °C	-30.063 °C	-29.2 °C	0.863 °C	± 1.0 °C	± 4.0E-01 °C
-15 °C	-15.242 °C	-14.5 °C	0.742 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	0.100 °C	0.7 °C	0.600 °C	± 1.0 °C	± 4.0E-01 °C
10 °C	10.127 °C	10.7 °C	0.573 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.730 °C	15.2 °C	0.470 °C	± 0.5 °C	± 4.0E-01 °C