

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.201 °C	0.0 °C	0.201 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	4.913 °C	5.0 °C	0.087 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.127 °C	10.0 °C	-0.127 °C	± 0.45 °C	± 4.0E-01 °C