

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	9.787 °C	9.5 °C	-0.287 °C	± 0.45 °C	± 4.0E-01 °C
32 °C	31.927 °C	32.0 °C	0.073 °C	± 0.45 °C	± 4.0E-01 °C
35 %HR	34.537 %HR	39.0 %HR	4.463 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.059 %HR	72.0 %HR	-3.059 %HR	± 3.0 %HR	± 1.9E+00 %HR