

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.440 °C        | -25.7 °C  | -0.260 °C     | ± 0.8 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.121 °C         | 0.1 °C    | 0.221 °C      | ± 0.8 °C             | ± 4.0E-01 °C                |
| 8 °C              | 8.032 °C          | 8.2 °C    | 0.168 °C      | ± 0.8 °C             | ± 4.0E-01 °C                |