

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0010 in	0.0010 in	0.0000 in	± 0.0002 in	±	7.0E-02 in
0.0020 in	0.0020 in	0.0000 in	± 0.0002 in	±	5.7E-02 in
0.0030 in	0.0030 in	0.0000 in	± 0.0002 in	±	7.6E-02 in
0.0040 in	0.0040 in	0.0000 in	± 0.0002 in	±	9.4E-02 in
0.0050 in	0.0050 in	0.0000 in	± 0.0002 in	±	7.0E-02 in
0.0060 in	0.0060 in	0.0000 in	± 0.0002 in	±	7.6E-02 in
0.0080 in	0.0080 in	0.0000 in	± 0.0002 in	±	1.0E-01 in
0.0090 in	0.0090 in	0.0000 in	± 0.0002 in	±	1.0E-01 in
0.0100 in	0.0100 in	0.0000 in	± 0.0002 in	±	9.4E-02 in