

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.007 °C	10.1 °C	0.093 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.035 °C	20.0 °C	-0.035 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.163 °C	30.2 °C	0.037 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.291 °C	40.3 °C	0.009 °C	± 0.5 °C	± 4.0E-01 °C