

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
200.1 lx	193 lx	7.08 lx	± 20 lx	±	1.1E+02 lx
1,000.9 lx	1,028 lx	-27.09 lx	± 60 lx	±	5.9E+02 lx
5,012.0 lx	5,060 lx	-47.97 lx	± 160 lx	±	2.9E+03 lx
15,035.7 lx	14,970 lx	65.66 lx	± 851 lx	±	8.6E+03 lx
20,098.7 lx	20,200 lx	-101.35 lx	± 1104.9 lx	±	1.2E+04 lx
50,280.6 lx	50,000 lx	280.61 lx	± 2514 lx	±	2.9E+04 lx