

Certificado No. (Certificate #):		58580		Equipo (Instrument):		Reloj Indicador			
Fecha de Calibración (Calibration Date):		2025-07-04		Marca (Brand):		-			
Patrón (Standard)		UBP (UUT)		Error (Error)		Exactitud (Accuracy)		Incertidumbre (Uncertainty)	
0.0500 in		0.050 in		0.000 in		± in		±	5.8E-04 in
0.1000 in		0.100 in		0.000 in		± in		±	5.8E-04 in
0.1050 in		0.105 in		0.000 in		± in		±	5.8E-04 in
0.1500 in		0.150 in		0.000 in		± in		±	5.9E-04 in
0.2000 in		0.200 in		0.000 in		± in		±	5.9E-04 in
0.3000 in		0.300 in		0.000 in		± in		±	6.1E-04 in
0.4000 in		0.400 in		0.000 in		± in		±	6.4E-04 in
0.5000 in		0.500 in		0.000 in		± in		±	6.8E-04 in
1.0000 in		1.000 in		0.000 in		± in		±	9.1E-04 in