

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
8 °C	8.028 °C	8.32 °C	0.292 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.891 °C	25.10 °C	0.213 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.807 °C	30.02 °C	0.213 °C	± 0.5 °C	± 4.0E-01 °C
70 %HR	70.478 %HR	69.46 %HR	-1.022 %HR	± 3.0 %HR	± 1.9E+00 %HR