

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
8 °C	7.916 °C	8.26 °C	0.340 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.878 °C	25.11 °C	0.229 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.832 °C	30.00 °C	0.172 °C	± 0.5 °C	± 4.0E-01 °C
70 %HR	70.466 %HR	69.67 %HR	-0.800 %HR	± 3.0 %HR	± 1.9E+00 %HR