

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 0 °C              | 0.331 °C          | 0.4 °C    | 0.069 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 5.144 °C          | 5.2 °C    | 0.056 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.458 °C         | 10.6 °C   | 0.142 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.366 %HR        | 28.6 %HR  | -1.766 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 50.069 %HR        | 52.8 %HR  | 2.731 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.243 %HR        | 69.3 %HR  | -0.943 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |