

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.414 °C	12.4 °C	-0.014 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.054 °C	26.8 °C	-0.254 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.146 °C	42.1 °C	-0.046 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.288 %HR	53.7 %HR	1.412 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.135 %HR	73.1 %HR	0.965 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.339 %HR	94.9 %HR	2.561 %HR	± 3.0 %HR	± 1.9E+00 %HR