

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.3 μm	23.4 μm	0.1 μm	± 1.03 μm	±	5.8E-02 μm
51.6 μm	51.4 μm	-0.2 μm	± 1.3 μm	±	5.8E-02 μm
124.6 μm	124 μm	-0.6 μm	± 2.0 μm	±	5.8E-02 μm
250.3 μm	248 μm	-2.3 μm	± 3.3 μm	±	5.8E-02 μm
486.0 μm	483 μm	-3.0 μm	± 5.7 μm	±	5.8E-02 μm
963.0 μm	961 μm	-2.0 μm	± 10 μm	±	9.6E-02 μm