

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.331 °C	0.3 °C	-0.031 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.144 °C	5.3 °C	0.156 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.458 °C	10.8 °C	0.342 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	32.9 %HR	2.534 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.069 %HR	48.2 %HR	-1.869 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	67.5 %HR	-2.743 %HR	± 3.0 %HR	± 1.9E+00 %HR