

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.331 °C	0.5 °C	0.169 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.144 °C	5.2 °C	0.056 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.458 °C	10.6 °C	0.142 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	31.5 %HR	1.134 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.069 %HR	51.5 %HR	1.431 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68.8 %HR	-1.443 %HR	± 3.0 %HR	± 1.9E+00 %HR