

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.331 °C	0.4 °C	0.069 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.144 °C	5.2 °C	0.056 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.458 °C	10.6 °C	0.142 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	32.6 %HR	2.234 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.069 %HR	48.7 %HR	-1.369 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	67.7 %HR	-2.543 %HR	± 3.0 %HR	± 1.9E+00 %HR