

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
12 °C	12.414 °C	12.0 °C	-0.414 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.054 °C	27.2 °C	0.146 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.146 °C	42.1 °C	-0.046 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.288 %HR	53.6 %HR	1.312 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.135 %HR	74.7 %HR	2.565 %HR	± 3.0 %HR	± 1.9E+00 %HR
92 %HR	92.339 %HR	93.2 %HR	0.861 %HR	± 3.0 %HR	± 1.9E+00 %HR