

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.414 °C	12.2 °C	-0.214 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.054 °C	27.2 °C	0.146 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.146 °C	42.5 °C	0.354 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.288 %HR	51.9 %HR	-0.388 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.135 %HR	72.2 %HR	0.065 %HR	± 3.0 %HR	± 1.9E+00 %HR
92 %HR	92.339 %HR	91.1 %HR	-1.239 %HR	± 3.0 %HR	± 1.9E+00 %HR