

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
12 °C	12.414 °C	12.5 °C	0.086 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.054 °C	27.5 °C	0.446 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.146 °C	42.5 °C	0.354 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.288 %HR	53.5 %HR	1.212 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.135 %HR	73.5 %HR	1.365 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.339 %HR	89.5 %HR	-2.839 %HR	± 3.0 %HR	± 1.9E+00 %HR