

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	20.586 °C	20.8 °C	0.214 °C	± 0.6 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	28.3 %HR	-2.270 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.642 %HR	51.7 %HR	1.058 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	78.6 %HR	3.040 %HR	± 3.0 %HR	± 1.9E+00 %HR