

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.985 °C	2.3 °C	0.315 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	4.7 °C	-0.293 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.002 °C	7.7 °C	-0.302 °C	± 0.5 °C	± 4.0E-01 °C