

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
3600.105 s	3600 s	-0.105 s	-0.003%	± 0.01 %	±	1.0E+03 s
3600.011 s	3600 s	-0.011 s	0.000%	± 0.01 %	±	1.0E+03 s
3600.094 s	3600 s	-0.094 s	-0.003%	± 0.01 %	±	1.0E+03 s