

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09920	0.800	± 1.2	± 4.3E-01 mm					
0 a 0.5	0.49908	0.920	± 1.6	± 4.3E-01 mm					
0 a 1	0.99917	0.830	± 2	± 4.3E-01 mm					
0 a 2	1.99910	0.900	± 2.8	± 4.8E-01 mm					
0 a 3	2.99914	0.860	± 3.6	± 4.8E-01 mm					
0 a 4	3.99924	0.760	± 4.4	± 4.8E-01 mm					
0 a 5	4.99939	0.610	± 5.2	± 4.8E-01 mm					
0 a 6	5.99894	1.060	± 6	± 3.6E-01 mm					
0 a 7	6.99960	0.400	± 6.8	± 3.6E-01 mm					
0 a 8	7.99917	0.830	± 7.6	± 3.6E-01 mm					