

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 30 °C                      | 30.614 °C                  | 30.7 °C            | 0.086 °C               | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 50 °C                      | 50.519 °C                  | 51.0 °C            | 0.481 °C               | ± 1.0 °C                      | ± 1.1E+00 °C                         |
| 35 %HR                     | 35.580 %HR                 | 38.0 %HR           | 2.420 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 75 %HR                     | 75.560 %HR                 | 76.0 %HR           | 0.440 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |