

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.614 °C	30.5 °C	-0.114 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.519 °C	50.9 °C	0.381 °C	± 1.0 °C	± 1.1E+00 °C
35 %HR	35.580 %HR	37.0 %HR	1.420 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	73.0 %HR	-2.560 %HR	± 5.0 %HR	± 1.9E+00 %HR