

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
8 °C	7.707 °C	7.15 °C	-0.554 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.946 °C	25.15 °C	0.207 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.045 °C	29.98 °C	-0.061 °C	± 0.5 °C	± 4.0E-01 °C
70 %HR	68.333 %HR	69.75 %HR	1.420 %HR	± 3.0 %HR	± 1.9E+00 %HR