

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 8 °C                       | 7.636 °C                   | 8.04 °C            | 0.400 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 24.712 °C                  | 24.13 °C           | -0.587 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 °C                      | 29.767 °C                  | 29.93 °C           | 0.165 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 70 %HR                     | 67.427 %HR                 | 68.75 %HR          | 1.326 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |