

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.614 °C	30.4 °C	-0.214 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	31.7 %HR	1.130 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.642 %HR	50.5 %HR	-0.142 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	67.6 %HR	-2.745 %HR	± 3.0 %HR	± 1.9E+00 %HR