

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.614 °C	30.5 °C	-0.114 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	28.8 %HR	-1.770 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.642 %HR	50.1 %HR	-0.542 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	69.5 %HR	-0.845 %HR	± 3.0 %HR	± 1.9E+00 %HR