

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
ft	ft	in	in	(Uncertainty)	ft	ft	in	in	(Uncertainty)
0 a 5	5.00182	-0.022	± 0.26	± 1.4E-03 in					
0 a 10	10.00241	-0.029	± 0.26	± 1.4E-03 in					
0 a 15	15.00165	-0.020	± 0.26	± 1.4E-03 in					
0 a 20	20.00141	-0.017	± 0.26	± 1.4E-03 in					
0 a 25	25.00163	-0.020	± 0.26	± 1.4E-03 in					
0 a 30	30.00117	-0.014	± 0.26	± 1.4E-03 in					
0 a 35	35.00141	-0.017	± 0.26	± 1.4E-03 in					
0 a 40	40.00129	-0.016	± 0.26	± 1.4E-03 in					
0 a 45	45.00234	-0.028	± 0.26	± 1.4E-03 in					
0 a 50	50.00258	-0.031	± 0.26	± 1.4E-03 in					