

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
ft	ft	in	in	(Uncertainty)	ft	ft	in	in	(Uncertainty)
0 a 5	5.00210	-0.025	± 0.26	± 1.4E-03 in					
0 a 10	10.00224	-0.027	± 0.26	± 1.4E-03 in					
0 a 15	15.00170	-0.020	± 0.26	± 1.4E-03 in					
0 a 20	20.00183	-0.022	± 0.26	± 1.4E-03 in					
0 a 25	25.00146	-0.018	± 0.26	± 1.4E-03 in					
0 a 30	30.00153	-0.018	± 0.26	± 1.4E-03 in					
0 a 35	35.00160	-0.019	± 0.26	± 1.4E-03 in					
0 a 40	40.00169	-0.020	± 0.26	± 1.4E-03 in					
0 a 45	45.00179	-0.022	± 0.26	± 1.4E-03 in					
0 a 50	50.00171	-0.021	± 0.26	± 1.4E-03 in					