

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
ft	ft	in	in	(Uncertainty)	ft	ft	in	in	(Uncertainty)
0 a 5	5.00200	-0.024	± 0.26	± 1.4E-03 in					
0 a 10	10.00226	-0.027	± 0.26	± 1.4E-03 in					
0 a 15	15.00122	-0.015	± 0.26	± 1.4E-03 in					
0 a 20	20.00160	-0.019	± 0.26	± 1.4E-03 in					
0 a 25	25.00145	-0.017	± 0.26	± 1.4E-03 in					
0 a 30	30.00138	-0.017	± 0.26	± 1.4E-03 in					
0 a 35	35.00130	-0.016	± 0.26	± 1.4E-03 in					
0 a 40	40.00145	-0.017	± 0.26	± 1.4E-03 in					
0 a 45	45.00159	-0.019	± 0.26	± 1.4E-03 in					
0 a 50	50.00128	-0.015	± 0.26	± 1.4E-03 in					