

Tipo de Medición <i>(Measurement Type)</i>	Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
Throat Set	3 mm	3 mm	0 mm	± 10 µm	±	5.8E+02 µm
Throat Set	11 mm	11 mm	0 mm	± 10 µm	±	5.8E+02 µm
Throat Set	15 mm	15 mm	0 mm	± 10 µm	±	5.8E+02 µm
Escala Movil	2 mm	2 mm	0 mm	± 10 µm	±	5.8E+02 µm
Escala Movil	8 mm	8 mm	0 mm	± 10 µm	±	5.8E+02 µm
Escala Movil	16 mm	16 mm	0 mm	± 10 µm	±	5.8E+02 µm
Escala Movil	26 mm	26 mm	0 mm	± 10 µm	±	5.8E+02 µm