

Tipo de Medición <i>(Measurement Type)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
Theoretical Throat	5 mm	5 mm	0 mm	± 1 µm	±	2.9E+02 µm
Theoretical Throat	15 mm	15 mm	0 mm	± 1 µm	±	2.9E+02 µm
Theoretical Throat	0.25 in	0.250 in	0 in	± 1 µm	±	2.9E+02 µm
Theoretical Throat	0.5 in	0.5 in	0 in	± 1 µm	±	2.9E+02 µm
Regla	20 mm	20 mm	0 mm	± 1 µm	±	2.9E+02 µm
Regla	50 mm	50 mm	0 mm	± 1 µm	±	2.9E+02 µm
Regla	1 in	1 in	0 in	± 1 µm	±	2.9E+02 µm
Regla	2 in	2 in	0 in	± 1 µm	±	2.9E+02 µm
Undercut	20 mm	20 mm	0 mm	± 1 µm	±	2.9E+02 µm
Undercut	1 in	1 in	0 mm	± 1 µm	±	2.90E+02 µm