

Tipo de Medición <i>(Measurement Type)</i>	Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
Fillet Weld Leg	5 mm	5 mm	0 mm	± 10 µm	±	5.8E-01 µm
Fillet Weld Leg	10 mm	10 mm	0 mm	± 10 µm	±	5.8E-01 µm
Fillet Weld Leg	20 mm	20 mm	0 mm	± 10 µm	±	5.8E-01 µm
Max Convexity	5 mm	5 mm	0 mm	± 10 µm	±	5.8E-01 µm
Max Convexity	10 mm	10 mm	0 mm	± 10 µm	±	5.8E-01 µm
Max Convexity	20 mm	20 mm	0 mm	± 10 µm	±	5.8E-01 µm
Max Concavity	5 mm	5 mm	0 mm	± 10 µm	±	5.8E-01 µm
Max Concavity	10 mm	10 mm	0 mm	± 10 µm	±	5.8E-01 µm
Max Concavity	20 mm	20 mm	0 mm	± 10 µm	±	5.8E-01 µm