

Certificado No. (Certificate #):		58604		Equipo (Instrument):		Medidor de Vibración			
Fecha de Calibración (Calibration Date):		2025-05-14		Marca (Brand):		Adash			
Patrón (Standard)		UBP (UUT)		Error (Error)		Exactitud (Accuracy)		Incertidumbre (Uncertainty)	
7 Hz	0.283 g-rms	0.28 g-rms	0.00 g-rms	± 0.0085 g-rms	±	1.2E-02 g-rms			
10 Hz	0.566 g-rms	0.56 g-rms	-0.01 g-rms	± 0.0170 g-rms	±	1.5E-02 g-rms			
50 Hz	0.707 g-rms	0.70 g-rms	-0.01 g-rms	± 0.0212 g-rms	±	1.7E-02 g-rms			
100 Hz	0.707 g-rms	0.69 g-rms	-0.02 g-rms	± 0.0212 g-rms	±	1.8E-02 g-rms			
1000 Hz	0.707 g-rms	0.72 g-rms	0.01 g-rms	± 0.0212 g-rms	±	1.8E-02 g-rms			
5000 Hz	0.707 g-rms	0.71 g-rms	0.00 g-rms	± 0.0212 g-rms	±	2.4E-02 g-rms			