

Certificado No. (Certificate #):		58616		Equipo (Instrument):		Reloj Indicador			
Fecha de Calibración (Calibration Date):		2025-07-03		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		Exactitud (Accuracy)		Incertidumbre (Uncertainty)	
1.0000 mm		0.99 mm		-8 µm		± 20 µm		± 5.8E+00 µm	
1.1000 mm		1.09 mm		-7 µm		± 20 µm		± 5.8E+00 µm	
1.5000 mm		1.50 mm		-2 µm		± 20 µm		± 5.8E+00 µm	
2.0000 mm		2.00 mm		0 µm		± 20 µm		± 5.8E+00 µm	
5.0000 mm		4.99 mm		-8 µm		± 20 µm		± 5.8E+00 µm	
10.0000 mm		9.99 mm		-10 µm		± 20 µm		± 5.8E+00 µm	
15.0000 mm		15.00 mm		-2 µm		± 20 µm		± 5.8E+00 µm	
20.0000 mm		20.00 mm		-3 µm		± 20 µm		± 5.8E+00 µm	
25.0000 mm		25.01 mm		10 µm		± 20 µm		± 5.8E+00 µm	