

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.000 mm	0.98 mm	-20 µm	± 20 µm	±	5.8E+00 µm
1.250 mm	1.24 mm	-10 µm	± 20 µm	±	5.8E+00 µm
1.500 mm	1.48 mm	-20 µm	± 20 µm	±	5.8E+00 µm
2.000 mm	1.98 mm	-20 µm	± 20 µm	±	5.8E+00 µm
3.000 mm	2.99 mm	-10 µm	± 20 µm	±	5.8E+00 µm
5.000 mm	4.98 mm	-20 µm	± 20 µm	±	5.8E+00 µm
8.000 mm	7.98 mm	-20 µm	± 20 µm	±	5.8E+00 µm
10.000 mm	9.99 mm	-10 µm	± 20 µm	±	5.8E+00 µm
12.000 mm	11.98 mm	-20 µm	± 20 µm	±	5.8E+00 µm