

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1 mm	0.99 mm	-10 µm	± 20 µm	±	5.8E+00 µm
1.25 mm	1.25 mm	0 µm	± 20 µm	±	5.8E+00 µm
1.5 mm	1.51 mm	10 µm	± 20 µm	±	5.8E+00 µm
2 mm	2.00 mm	0 µm	± 20 µm	±	5.8E+00 µm
3 mm	3.01 mm	10 µm	± 20 µm	±	5.8E+00 µm
5 mm	5.01 mm	10 µm	± 20 µm	±	5.8E+00 µm
8 mm	8.01 mm	10 µm	± 20 µm	±	5.8E+00 µm
10 mm	10.00 mm	0 µm	± 20 µm	±	5.8E+00 µm
12 mm	12.01 mm	10 µm	± 20 µm	±	5.8E+00 µm