

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1 mm	1.00 mm	0 µm	± 20 µm	±	5.8E+00 µm
2 mm	2.01 mm	10 µm	± 20 µm	±	5.8E+00 µm
3 mm	3.00 mm	0 µm	± 20 µm	±	5.8E+00 µm
5 mm	5.00 mm	0 µm	± 20 µm	±	5.8E+00 µm
8 mm	8.01 mm	10 µm	± 20 µm	±	5.8E+00 µm
10 mm	10.01 mm	10 µm	± 20 µm	±	5.8E+00 µm
15 mm	15.01 mm	10 µm	± 20 µm	±	5.8E+00 µm
20 mm	20.00 mm	0 µm	± 20 µm	±	5.8E+00 µm
25 mm	25.00 mm	0 µm	± 20 µm	±	5.8E+00 µm