

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | E.M.P <i>(Tolerance)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|--------------------------|------------------------------------|------------|
| 23.3 μm                  | 24.0 μm          | 0.70 μm              | ± 3 μm                   | ±                                  | 5.8E-01 μm |
| 51.6 μm                  | 53.0 μm          | 1.40 μm              | ± 3 μm                   | ±                                  | 5.8E-01 μm |
| 124.6 μm                 | 126.0 μm         | 1.40 μm              | ± 3 μm                   | ±                                  | 5.9E-01 μm |
| 250.3 μm                 | 252.0 μm         | 1.70 μm              | ± 3 μm                   | ±                                  | 6.3E-01 μm |
| 486.00 μm                | 485.0 μm         | -1.00 μm             | ± 3 μm                   | ±                                  | 7.5E-01 μm |
| 963.00 μm                | 962.0 μm         | -1.00 μm             | ± 3 μm                   | ±                                  | 1.1E+00 μm |