

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 30 °C                      | 30.564 °C                  | 30.5 °C            | -0.064 °C              | ± 0.6 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 30.263 %HR                 | 30.0 %HR           | -0.263 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.028 %HR                 | 48.7 %HR           | -1.328 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.294 %HR                 | 68.0 %HR           | -2.294 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |