

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.564 °C	31.1 °C	0.536 °C	± 0.6 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	30.5 %HR	0.237 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	48.4 %HR	-1.628 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.294 %HR	67.3 %HR	-2.994 %HR	± 3.0 %HR	± 1.9E+00 %HR