

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.564 °C	30.6 °C	0.036 °C	± 0.6 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	31.8 %HR	1.537 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	50.6 %HR	0.572 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.294 %HR	68.4 %HR	-1.894 %HR	± 3.0 %HR	± 1.9E+00 %HR