

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.089 °C        | -24.8 °C  | 0.289 °C      | ± 0.95 °C            | ± 4.0E-01 °C                |
| 0 °C              | 0.310 °C          | 0.5 °C    | 0.190 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 5 °C              | 5.194 °C          | 5.5 °C    | 0.306 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 10 °C             | 10.107 °C         | 10.4 °C   | 0.293 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 25 °C             | 25.249 °C         | 25.6 °C   | 0.351 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 30 °C             | 30.163 °C         | 30.3 °C   | 0.137 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 40 °C             | 40.641 °C         | 41.0 °C   | 0.359 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |

