

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
# 0 mm	0.000 mm	0 µm	± 4 µm	±	5.8E+00 µm
1.0 mm	1.001 mm	1 µm	± 4 µm	±	5.8E+00 µm
1.25 mm	1.251 mm	0 µm	± 4 µm	±	5.8E+00 µm
1.5 mm	1.500 mm	0 µm	± 4 µm	±	5.8E+00 µm
2 mm	2.000 mm	0 µm	± 4 µm	±	5.8E+00 µm
3 mm	3.000 mm	0 µm	± 4 µm	±	5.8E+00 µm
5 mm	5.001 mm	1 µm	± 4 µm	±	5.8E+00 µm
7 mm	7.000 mm	0 µm	± 4 µm	±	5.8E+00 µm
10 mm	10.000 mm	0 µm	± 4 µm	±	5.8E+00 µm
15 mm	15.000 mm	0 µm	± 4 µm	±	5.8E+00 µm
25 mm	25.000 mm	0 µm	± 4 µm	±	5.8E+00 µm
50 mm	50.000 mm	0 µm	± 4 µm	±	5.8E+00 µm